



Labeling Insert for Bioquell Hydrogen Peroxide Sterilant EPA Registration Number 72372-1-86703



1	OVERVIEW	3
2	HPV BIO-DECONTAMINATION	3
2.1	CONDITIONING	3
2.2	GASSING	3
2.3	DWELL	4
2.4	AERATION	4
3	USER SAFETY REQUIREMENTS	4
3.1	RESPIRATOR USE	4
3.2	HANDLING BIOQUELL HYDROGEN PEROXIDE STERILANT	4
4	EFFICACY	6
5	BIO-DECONTAMINATION CYCLE PROTOCOL, (BCP)	6
5.1	PREPARATION OF THE BCP - STEP 1: PRE-CYCLE PLANNING	7
5.1.1	Target Area Dossier	7
5.1.2	Pre-cycle Planning File	9
5.2	STEP 2: NOTIFICATION	10
5.2.1	Personnel Briefing	10
5.2.2	Cycle Operator Briefing	10
5.2.3	Signage	10
5.3	STEP 3: TARGET ENCLOSURE SEALING	10
5.4	STEP 4: TARGET ENCLOSURE PREPARATION	11
5.4.1	Cleaning	11
5.4.2	Absorbent Materials	11
5.4.3	Occluded Surfaces	11
5.4.4	Extremes of Temperature	11
5.4.5	Active Air-paths	12
5.5	STEP 5: CYCLE START	12
5.6	STEP 6: MONITORING	12
5.6.1	Leak Monitoring	12
5.6.2	Cycle Monitoring	13
5.7	STEP 7: CYCLE COMPLETION	13
5.7.1	Cycle Finish Verification	13
5.7.2	Post-cycle Re-entry	13
5.7.3	Cycle Success Criteria	13
6	VALIDATED AND NON-VALIDATED USE	14
6.1	VALIDATED USE IN ENCLOSURES UP TO 3,500FT ³	14
6.2	VALIDATED USE IN ENCLOSURES GREATER THAN 3,500FT ³	14
6.2.1	Biological Indicators, BIs	15
6.2.2	Chemical Indicators, CIs	16



1 Overview

Bioquell Hydrogen Peroxide Sterilant has been registered by Bioquell in accordance with the Federal regulations for use in accordance with the instructions listed in this document. The contents may only be used with Bioquell products in line with their user manuals and must not be used for any purpose other than that described.

Before using Bioquell Hydrogen Peroxide Sterilant the operator should ensure that he / she has undergone appropriate training on the specific Bioquell HPV generator and has been certified as such. If unsure, refresher training should be arranged before using the unit to run a bio-decontamination cycle.

2 HPV Bio-decontamination

When bio-decontaminating an enclosure (the “Enclosure”) using hydrogen peroxide vapor (“HPV”), the operator uses the Bioquell HPV generator to inject HPV into the atmosphere of the Enclosure resulting, once saturation conditions have been reached, with the formation of a very thin layer of ‘micro-condensation’ onto every exposed surface within the Enclosure. It is the formation of this microscopic layer of hydrogen peroxide condensate that provides the rapid efficacy of the bio-decontamination process and thus the success of the bio-decontamination cycle itself.

Bioquell’s unique HPV generation technology does not pressurize the Enclosure and thus (with adequate area sealing and monitoring) can be safely deployed across a range of different chambers, rooms or enclosures including those facilities that have not been purpose built and designed for use with gaseous sterilants.

Upon completion of the active phase of the bio-decontamination cycle the HPV is catalytically converted into oxygen and water vapor (humidity), thus negating the need for post injection extraction of the vapor via the ventilation system.

A typical hydrogen peroxide vapor bio-decontamination cycle is made up of 4 distinct phases, each of which is described below.

2.1 Conditioning

The conditioning phase is made up of internal system tests within the unit along with the heating of the vaporizer in preparation for the start of the gassing cycle.

2.2 Gassing

During the gassing phase the Bioquell HPV generator flash evaporates the Bioquell Hydrogen Peroxide Sterilant to generate HPV which is then injected into an air-stream at a specified rate. The active distribution system injects the HPV into the sealed target enclosure resulting in an increase in the concentration of HPV and, at saturation, producing micro-condensation deposition onto surfaces.



2.3 Dwell

Following the completion of the gassing phase, a pre-established, timed dwell phase results in the HPV circulating throughout the Enclosure ensuring that the HPV sterilant has sufficient contact time with the biological agents to ensure a successful bio-decontamination.

2.4 Aeration

The aeration phase results in the removal of the HPV from the target area, reducing the vapor concentration to < 1PPM, the OSHA PEL (Permitted Exposure Limit) in the United States. This is typically achieved by the catalytic conversion of the HPV into water vapor and oxygen.

The aeration phase can be accelerated using proprietary Bioquell equipment, designed and tested for this purpose and application. The process can be further accelerated, where appropriate, using the ventilation system serving the Enclosure. However, use of the ventilation to purge any gaseous sterilant must only be conducted by trained personnel. The restricted access status of the Enclosure may only be revoked upon completion of the aeration phase when the vapor concentration has been verified throughout as < 1PPM or local equivalent exposure level.

3 User Safety Requirements

3.1 Respirator use

In the event of a respirator being required for use with this product, the trained operator conducting the bio-decontamination shall ensure that:

- All respirators are fit for purpose and are fitted with suitable NIOSH approved filters suitable for the perceived breathing hazards.
- All respirator users have undergone appropriate respirator training and have been checked to ensure that they are medically able to wear, and work in the respirator safely and comfortably.

3.2 Handling Bioquell Hydrogen Peroxide Sterilant

Bioquell Hydrogen Peroxide Sterilant contains the active ingredient hydrogen peroxide. Liquid hydrogen peroxide is classified as corrosive and must be handled with the utmost care and whilst wearing appropriate personnel protection equipment, ("PPE"). After handling, users should remove all PPE immediately and wash their hands before eating, drinking, or using the bathroom. Hydrogen peroxide vapor is also harmful in high concentrations and as such liquid hydrogen peroxide should only be handled in open areas or those that have adequate ventilation.

A summary of the health and safety information concerning liquid hydrogen peroxide is shown below, and any PPE used when handling liquid hydrogen peroxide that is not disposable must be maintained in accordance with manufacturers' recommendations.

<p>Skin</p> 	<p>Contact with the skin causes irritation and possible burns. May also cause discoloration, swelling and the formation of blisters. Prolonged or repeated contact may cause dermatitis.</p> <p>Drench the skin thoroughly with water (and wash with soap if available). If spilled on clothing, remove immediately and wash before reuse. Skin contact may cause discoloration (whiteness) of the skin, but this is only temporary and the skin will quickly recover its normal color, usually within one hour. If symptoms persist, seek medical advice.</p>
<p>Eyes</p> 	<p>Contact is potentially very serious. Eye contact is corrosive and may cause severe burns, corneal damage and blindness.</p> <p>Do NOT allow victim to rub or keep eyes closed. Irrigate with an eye wash kit or water for at least 10 minutes whilst holding the eyelids open. Seek immediate medical advice.</p>
<p>Mouth / Ingestion</p> 	<p>Corrosive and irritating to the mouth, throat, and abdomen. Large doses may cause symptoms of abdominal pain, vomiting, and diarrhea as well as blistering or tissue destruction. Stomach distensions (due to rapid liberation of oxygen), and risk of stomach perforation, convulsions, fluid on the lungs or brain, coma, and death are possible.</p> <p>Do NOT induce vomiting. Rinse mouth thoroughly with water and give the casualty plenty of water to drink. Seek immediate medical advice.</p>
<p>Vapor</p> 	<p>Hydrogen peroxide vapor can cause irritation to the eyes, nose, throat, lungs and skin. Acute damage to the respiratory tract may also occur at high concentrations. More serious consequences include insomnia, nervous tremors with numb extremities, chemical pneumonia, unconsciousness and death.</p> <p>Remove casualty immediately to fresh air, rest and keep warm. Seek immediate medical advice.</p>
<p>Fire</p> 	<p>During a fire, highly toxic gases may be generated by thermal decomposition. Do not attempt to tackle a hydrogen peroxide fire. Call the fire department and ask for chemical emergency team. (Water only should be used on a hydrogen peroxide fire).</p>



4 Efficacy

Bioquell Hydrogen Peroxide Sterilant is only to be used with Bioquell HPV generators and when used correctly is a highly effective sterilant, active against spores, bacteria, viruses, and fungi on exposed, pre-cleaned dry porous and non-porous surfaces in sealed enclosures. Bioquell solution can be used in healthcare, pharmaceutical, defense, university and life sciences sectors including hospitals, office blocks, outdoor sheds, aircraft, retail facilities, restaurants, power stations, schools, factories, laboratories, marine craft, army vehicles, bars, sewage works, nursing homes, pharmaceutical buildings, warehouse/storage facilities, governmental buildings. Mobile medical facilities, domestic residences, water treatment plants.

When Bioquell solution is used in conjunction with Bioquell HPV generators, the following validated cycles shall apply.

For use as a sterilant, sporicide, fungicide, bactericide, and virucide in sealed enclosures up to 35ft³, inject 3.1g/min for 55 minutes, followed by a 180 minute dwell, followed by aeration.

For use as a sterilant, sporicide, fungicide, bactericide, and virucide in sealed enclosures up to 3,500ft³, inject 10g/min for 150 minutes, followed by a 15 minute dwell, followed by aeration.

For use as a sporicide, fungicide, bactericide, and virucide in sealed enclosures greater than 3,500ft³ in accordance with instructions in part 6.

This product is designed to be used in Bioquell HPV generation equipment and may not be used with any other equipment other than for which it was designed. Use of this product in any manner other than for which it was designed is strictly prohibited and may not produce the desired results. This product is not to be used as a terminal disinfectant or sterilant for the processing of critical or semi-critical medical devices.

5 Bio-decontamination Cycle Protocol, (BCP)

Prior to commencing a bio-decontamination cycle of any target enclosure the individual responsible for decontaminating the Enclosure (the “Cycle Manager”) must ensure that he / she has adequate and current training and in liaison with the appropriate parties (e.g. the building manager, or supervisor of the proposed Enclosure) a bio-decontamination protocol has been established. This should cover all aspects of the bio-decontamination cycle and may include, but not be limited to:

- Health and safety considerations;
 - monitoring points and frequency,
 - an evacuation plan,
 - any impact on existing evacuation plans (i.e. will isolation of the target enclosure impact on an active fire escape),
 - target area signage,
 - emergency procedures,
 - PPE required for pre-cycle entry,

- PPE required for post cycle entry.
- Practical considerations;
 - ventilation configuration within the target area,
 - power requirements,
 - access to the target area,
 - biological indicator regime, if any, and location plan,
 - equipment location plan,
 - sealing of the target area,
 - Any changes to the fabric of the building or ventilation system since the previous bio-decontamination cycle was performed.

The BCP should be comprehensive and may ultimately take the format of a checklist to ensure that every necessary task has been completed by the Cycle Manager. The BCP should relate to the Enclosure and be appropriately detailed, although the contents of the BCP will by necessity vary with the relative complexity of the Enclosure and the frequency of the bio-decontamination. The aim of the BCP is to ensure that each bio-decontamination cycle is run in a safe, considered and efficient manner - and may also form part of a validation process where consistency and repeatability are important.

As standard procedure, prior to undertaking a bio-decontamination cycle the Cycle Manager and any other operators should re-acquaint themselves with this packaging material, the user manual and any additional training materials supplied with the Bioquell HPV generator. These should be read in context with any existing BCPs that have been established for use within the Enclosure, and any applicable local or state laws.

For facilities that are undergoing HPV bio-decontamination for the first time a new BCP should be produced. Subsequent bio-decontaminations of the same facility may be conducted using an existing BCP. The following sections provide a template that a typical BCP may follow although it must be noted that each bio-decontamination and target facility are inherently different and, as such, this list is not exhaustive and each prospective cycle must be considered individually and will present its own points to address.

5.1 Preparation of the BCP - Step 1: Pre-cycle planning

5.1.1 Target Area Dossier

In the cycle planning phase the potential target facility must be inspected and a meeting with any appropriate parties arranged in advance of the cycle. Appropriate people to liaise with may include (but are not limited to):

- The responsible person for the target enclosure (e.g. building manager, unit manager, ward sister, matron etc.)
- Site security.
- Safety officer.
- Fire officer.
- Quality Control.



The outcome of the Enclosure inspection should allow compilation of a dossier containing such information as.

- Contact details for both the Cycle Manager and any other responsible persons.
- Target area dimensions, layout, and floor plans.
- A list of any areas not within the target enclosure to be designated as restricted access for the duration of the bio-decontamination cycle, e.g. adjacent plant rooms or technical areas.
- A description of the normal purpose of the prospective target area
- Reason for the bio-decontamination and information concerning the nature of any potential contamination present in the area and subsequent PPE requirements.
- Access and security clearance protocols.
- Details of any on-site inductions / training required
- Photographs of the target facility where appropriate.
- Access information for transfer of equipment into (pre-cycle) and out of (post cycle) the target enclosure. This is particularly pertinent in aseptic/GMP or containment applications.
- A description, and where possible/appropriate drawings of the ventilation system serving the target facility, and the status of the system throughout each stage of the cycle. In certain configurations heating, ventilation and air-conditioning, (HVAC), systems have the potential to reduce cycle times by accelerating the aeration phase, but more importantly represent a severe risk of vapor leakage to other areas and, as such, **MUST** be clearly understood before proceeding with any aspect of the bio-decontamination. The following points should be considered:
 - Regardless of the configuration of the HVAC system, it **must** be isolated for the duration of the 'gassing', and 'dwell' phases of the bio-decontamination cycle, either by shutting the system down or by appropriately sealing the inlet and outlet ducts.
 - If the target facility has been designed with gaseous fumigation in mind **and** an appropriate HVAC fumigation protocol is in place then the HVAC system may be re-instated during the aeration phase. Reinstating the HVAC system serves to purge the HPV from the target enclosure and thus accelerate the aeration phase, and ultimately reduce the total cycle time.
 - If the HVAC system serving the target area is common to other areas outside the target area, or if the HVAC system cannot be shutdown within the enclosure then both inlet and outlet ducts must be sealed and remain sealed for the duration of the cycle. The sealing material may only be removed once the cycle has been completed, i.e. the HPV concentration within the target area is <1PPM and safe for personnel re-entry.
 - If the cycle manager is in any doubt as the configuration of the HVAC system then both inlet and extract ducts must be sealed for the duration of the cycle.
- A marked floor plan of the area showing possible vapor leakage paths, areas to be sealed and where signage is to be displayed. Signage must be in the appropriate languages (e.g. English and Spanish) and should display the



- emergency contact information of the cycle manager and remain in place for the duration of bio-decontamination cycle. Potential leakage paths to be illustrated on the plan and sealed include, but are not limited to,
- Doors and all access points to the facility (windows, pass through hatches, inspection ports, ill-fitting pipe-work).
 - All potential access points to any floors above and below the target facility including chutes, electrical risers, chimneys, hatches, damaged/loose flooring, pipe-fittings.
 - Damaged, missing, or worn sections of false ceilings within the facility.
 - Any potentially damaged areas of the fabric of the target area itself.
- A global plan/sketch of the area surrounding the target enclosure showing evacuation routes and the location of emergency equipment (e.g. fire extinguishers, fire alarm 'break-glass' points, emergency shower/eye-wash stations, telephones).
 - An evacuation plan in the event of an emergency listing muster points and a list of appropriate emergency contact telephone numbers including:
 - Cycle manager.
 - Target area responsible person (e.g. Unit Manager / supervisor).
 - On-site emergency personnel (if applicable).
 - Local emergency services (Fire, ambulance, police, hospitals).

Whilst it is essential that all areas are independently assessed for suitability, if there are a number of identical enclosures, or enclosures which are representative of each other, it is not essential that a new or full BCP is completed for every decontamination. However, the Cycle Manager must ensure that all processes and procedures are carried out in accordance with a generic dossier, with any enclosure specific alterations adhered to.

5.1.2 Pre-cycle Planning File

On completion of the target area dossier it should be compiled along with all other pertinent documentation to form a single file. Documents that should be compiled include the following (where applicable).

- A HVAC system bio-decontamination protocol (if applicable).
- Bio-decontamination cycle protocol, BCP, documents from previous cycles within the target area.
- The training records of the cycle manager and all other operators of the Bioquell HPV equipment.
- User manuals and calibration records for the Bioquell HPV generators and pertinent ancillary equipment to be used during the fumigation (e.g. Handheld HPV sensors).

Each facility will have its own protocols and procedures which govern actions on site. The list presented here is a guide and should be used as the basis of the planning file.

5.2 Step 2: Notification

5.2.1 Personnel Briefing

Prior to commencing any HPV bio-decontamination cycle it is of the utmost importance that all personnel who may have access to the target facility are made aware of the process. All staff/personnel should be briefed in-terms of the logistical factors (cycle timings, areas designated out-of-bounds, restricted access areas, monitoring points) and how their normal working practices may be impacted for the cycle duration and, of course, the health and safety aspects of HPV bio-decontamination.

If appropriate a briefing session should be arranged with key personnel that may routinely have access to the target area and they should be made aware of relevant aspects of the bio-decontamination to be performed including:

- Proposed cycle timings and timescales.
- Target enclosure boundaries.
- Other areas that are to be designated restricted access zones.
- Monitoring points during the cycle (e.g. is access required to patient areas adjacent to the target enclosure?).
- Emergency procedures and evacuation routes.
- Any impact on existing emergency procedures (i.e. does the target area obscure an active fire escape route if so alternative arrangements must be made prior to the cycle start).
- A background of HPV and the bio-decontamination process.

When performing a bio-decontamination cycle within a public building where access protocols and restrictions are more difficult to enforce the briefing should be given to relevant facility staff and additional personnel drafted in as appropriate to 'police' doors and access points to the target/restricted areas for the duration of the cycle.

5.2.2 Cycle Operator Briefing

Prior to the cycle start the cycle operators should have a separate briefing in which all aspects of the BCP are discussed in order to ensure that all cycle personnel are familiar with the detail of the proposed bio-decontamination schedule.

5.2.3 Signage

Each access point to the target enclosure must display adequate signage to ensure that no person may gain unauthorized access to the target enclosure. Signs should always adhere to local legislation and requirements.

5.3 Step 3: Target Enclosure Sealing

All possible leakage paths that were identified within the target area dossier must be adequately sealed to prevent leakage of HPV from the target enclosure but also to prevent the ingress of fresh air into the target enclosure that can dilute the HPV and impinge on the efficacy of the cycle. A number of sealing methods may be used



including non-marking adhesive tape, dedicated blanking sheets, or if sealing ventilation ducts then mechanical gas-tight dampers may be used as appropriate. Alternative methods to those listed are available.

When sealing potential leakage paths, they should be sealed using a secure method that will ensure the seal will be maintained for the duration of the bio-decontamination cycle, this is especially pertinent when sealing inlet ventilation grills where pressure can occasionally build up.

5.4 Step 4: Target Enclosure Preparation

Prior to commencing any bio-decontamination cycle the target enclosure should be optimized in order to maximize the efficacy and achieve a rapid and consistent bio-decontamination. There are a number of steps to be taken and these are listed and discussed below.

5.4.1 Cleaning

Hydrogen peroxide vapor has limited penetrating power into dirt and other gross contamination and thus prior to commencing the bio-decontamination cycle the target area must be subject to a minimum level of cleaning to ensure that the target area is *visibly clean* – i.e. free from all gross contamination including dust, dirt, blood, faeces, animal feed. If large levels of dust or dirt are present upon commencing the cycle then viable micro-organisms may well be present below the gross contamination and could possibly survive the bio-decontamination process.

5.4.2 Absorbent Materials

Whilst absorbent materials can safely remain within the target area and be exposed to the bio-decontamination cycle (although subsequent desorption will extend the aeration process) in many situations it is favorable for these to be removed if appropriate.

Consumables (gloves, paper towels etc.) can remain within the target area although the risk of surface area occlusion becomes an issue if large amounts of consumables are present within a small area.

5.4.3 Occluded Surfaces

HPV is not freely penetrating through many materials; as such it is vitally important that the occurrence of occluded (i.e. covered) surfaces is minimized. This is achieved by opening all cupboards and drawers to expose the internal surfaces (and contents) along with any other equipment within the area that can be opened or configured such to ensure the maximum surface area is exposed.

5.4.4 Extremes of Temperature

The hydrogen peroxide vapor bio-decontamination process relies on saturation of the atmosphere of the sealed target area with vapor in order to form a layer of micro-condensation of hydrogen peroxide that in turn affects the bio-decontamination; as



such any factors that can affect the formation of the condensate layer must be controlled. Temperature gradients within the target area should be avoided as cooler surfaces will see the formation of micro-condensation sooner and more plentifully than warmer areas and as such areas within the same room may not be exposed to the same cycle. Failure to do so may potentially lead to reduced efficacy of the bio-decontamination cycle due to uneven vapor distribution throughout the target enclosure.

In order to prevent the formation of temperature gradients all equipment that operates outside ambient room temperature should be shut down in advance of the cycle and allowed to return to ambient temperature prior to the commencement of the bio-decontamination cycle. Such equipment includes hot rooms, autoclaves, incubators, refrigerators, freezers, cold rooms, ovens etc.

5.4.5 Active Air-paths

As with all gaseous sterilants airflows present within the target area can impinge on vapor distribution and thus the cycle itself. Any equipment that re-circulates air flow solely within the target area can be left running throughout the cycle although if there is any internal filtration within the equipment itself the cycle parameters should be changed as appropriate to counteract the absorption within the filters. E.g. recirculatory safety hoods within the target area may be left running although the cycle should be elongated to counteract absorption into the filters. Equipment such as computers and laptops should be left running to draw HPV through the unit itself, thus bio-decontaminating the internals of the machine.

5.5 Step 5: Cycle Start

Before commencing the bio-decontamination cycle the Cycle Manager should go through the BCP as a checklist acknowledging that all necessary steps have been completed ensuring the safety of the cycle.

The facility manager should also confirm that all personnel who work within the target facility and any personnel who may have cause to access the area (e.g. cleaning or security staff) have been notified about the cycle and all evacuation and emergency procedures.

Upon completion of the acknowledgement procedures the Cycle Manager may then begin the bio-decontamination cycle.

5.6 Step 6: Monitoring

Monitoring the bio-decontamination cycle takes two distinct phases, monitoring the perimeter of the target enclosure for vapor leakage, and monitoring the vapor concentration within the target enclosure to monitor the cycle progress, and ultimately to confirm the end of the cycle.

5.6.1 Leak Monitoring



The cycle operators should use a hand held hydrogen peroxide sensor in order to verify that there is no escape of vapor from the target enclosure, by monitoring the perimeter of the target area along with any areas that are served by a common ventilation system. Leak monitoring should continue through the gassing and dwell phases of the bio-decontamination cycle.

5.6.2 Cycle Monitoring

The progress of the bio-decontamination cycle itself should (where applicable) be monitored using remote sensory equipment placed within the target area. The sensors should be configured such that they provide real-time data of the cycle parameters within the target enclosure. This data should then be logged at regular intervals throughout the cycle to record the cycle progress. On completion of the gassing and dwell phases, as the cycle moves into aeration the sensors allow verification of the vapor concentration for post-cycle re-entry.

5.7 Step 7: Cycle Completion

5.7.1 Cycle Finish Verification

A bio-decontamination cycle is completed once the cycle is in the aeration phase and the vapor concentration is below the applicable local exposure limit for personnel re-entry without the need for any respiratory apparatus, (<1PPM). The vapor concentration should first be verified using the remote sensors (where applicable) and if they read < 1PPM (or other appropriate local exposure limit) then personnel may re-enter the target enclosure with a high-resolution hand held hydrogen peroxide monitor to perform an enclosure concentration measurement. If the concentration throughout the facility is measured at less than 1PPM the bio-decontamination cycle has been completed.

5.7.2 Post-cycle Re-entry

Prior to re-entering the target enclosure personnel must first ensure that the measured vapor concentration is <1PPM (or appropriate local limit). The Cycle Manager should liaise with the facility or unit manager prior to re-entry to take into account any gowning/PPE protocols that may be in place within the area. This is especially pertinent when performing bio-decontamination cycles within aseptic or containment scenarios.

Upon verification that the vapor concentration is safe for re-entry and removal of all bio-decontamination and sampling equipment from within the target enclosure, all signage should be removed, all sealing materials removed and the area 'released' for normal operation.

5.7.3 Cycle Success Criteria

A bio-decontamination cycle may be declared successful if the validation standards defined in the BCP have been satisfied, and the aeration phase has been completed with the vapor concentration within the target enclosure confirmed as < 1PPM.



6 Validated and Non-Validated Use

6.1 Validated Use in Enclosures up to 3,500ft³

Validated bio-decontamination cycles utilizing Bioquell solution and Bioquell HPV generators have been developed for use as a sterilant, sporicide, fungicide, bactericide, and virucide in sealed enclosures up to 35ft³, and 3,500ft³.

The cycle parameters are,

For use as a sterilant, sporicide, fungicide, bactericide, and virucide in sealed enclosures up to 35ft³, inject 3.1g/min for 55 minutes, followed by a 180 minute dwell, followed by aeration.

For use as a sterilant, sporicide, fungicide, bactericide, and virucide in sealed enclosures up to 3,500ft³, inject 10g/min for 150 minutes, followed by a 15 minute dwell, followed by aeration.

6.2 Validated Use in Enclosures greater than 3,500ft³

Bioquell hydrogen peroxide solution may also be used as a sterilant, sporicide, fungicide, bactericide, and virucide in sealed enclosures of various volumes (including enclosures greater than 3,500ft³) with the development of a customized validated bio-decontamination cycle.

The set-up and cycle management phases of customized cycles are identical to those for a validated cycle with regard to the preparation of the bio-decontamination cycle protocol, ("BCP"), and target area set-up and sealing procedures.

In order for a customized cycle to be effective it is vital that the Cycle Manager gives due consideration to global vapor distribution throughout the target facility in order to ensure uniform formation of micro-condensation. As such, due consideration must be given to the number and location of Bioquell HPV generators deployed during the cycle, and the appropriate use of oscillating distribution fans or other appropriate equipment to ensure good vapor distribution. In accordance with the procedures described above the positions of all equipment used within the bio-decontamination cycle should be recorded on a facility plan within the BCP.

When performing customized validated cycles the cycle must be capable of attaining the required bio-burden reduction (as specified in the BCP), and have appropriate use of pre-determined indicators to ensure that the specified level is reached throughout the target facility.

On completion of the target area set-up and sealing procedures (including indicator placement) (sections 5.1 to 5.4), the Cycle Manager can begin the cycle; the cycle itself will have the same structure as a validated cycle with discrete conditioning, gassing, dwell and aeration phases.

Upon successful completion of the 'conditioning' phase (including system test) the cycle moves into the 'gassing' phase with HPV injected into the enclosure. The Cycle



Manager should monitor the cycle environmental data from within the target enclosure recorded via the on-board sensory equipment in order to recognize the point of onset of micro-condensation, the dew-point. Once micro-condensation has been achieved within the enclosure the cycle then moves into the 'dwell' phase in which the vapor is allowed to circulate within the target enclosure and ensure adequate contact time is allowed between the sterilant and the biological agents to affect a successful bio-decontamination.

Upon completion of the dwell phase the cycle moves in the aeration phase removing the HPV from the target area, reducing the vapor concentration to < 1PPM, the OSHA PEL (Permitted Exposure Limit). Once the vapor concentration has been confirmed as <1PPM, the restricted access status of the target facility may be revoked and the facility 'released' back into normal operation.

Should a cycle fail to meet the pre-determined target challenge then the cycle has not been successful and the cycle should be repeated with the gassing and/or dwell periods increased, and the validation process repeated.

When conducting any bio-decontamination cycle validated or non-validated all user safety procedures listed in section 3 and operational procedures in section 5 (including monitoring and post cycle re-entry) must be adhered to and overseen by the Cycle Manager.

6.2.1 Biological Indicators, BIs

In order to assess the success of bio-decontamination cycles a standard challenge is used to ensure that the cycle has been effective. Whilst various validation methods can be used biological indicators, (BIs), are the industry standard method for validation of hydrogen peroxide bio-decontamination cycles as they present the most consistent, and repeatable challenge.

A number of organisms may be used although the accepted organism is *Geobacillus stearothermophilus*; Bacillus endospores are the most resistant class of organisms to deactivation and thus provide suitable challenge organisms. *Geobacillus stearothermophilus* also has inherent practical operational advantages in that it is thermophilic with an optimum incubation temperature of 57°C, limiting the possibility of false positives due to the high incubation temperature. It is also a category 1 organism so is not harmful to humans and thus may be easily and safely handled.

The industrially accepted biological indicator challenge is a 6-log (i.e. > 1,000,000 spores per indicator) inoculum of *Geobacillus stearothermophilus* although lower challenge BIs (i.e. 4-log – 10,000 spores) are also commercially available. Experience has shown that the most consistent BIs are those that are inoculated onto a stainless steel substrate; other inoculum substrates including paper are available but experience has shown them to be less consistent and repeatable.

BIs should be placed throughout the target enclosure typically placed in the corners of rooms where a 'dead spot' in terms of vapor distribution is formed at the point where three walls meet. The number of indicators used is at the discretion of the cycle



manager but typically 1 BIs per 15m³, and each location should be recorded on a floor plan of the target enclosure and should be kept with the bio-decontamination plan.

Upon completion of the bio-decontamination cycle the BIs should be retrieved and incubated as per the organism protocols and the results available after the defined incubation period.

6.2.2 Chemical Indicators, CIs

Chemical indicators, (CIs) that change color in the presence of hydrogen peroxide vapor are also commercially available and may be used in cycle verification if calibrated to an appropriate biological challenge.